

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Nagalur  
 Lab No 12629  
 Month August  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	177 mg/dL	0.16	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11	1.8 mg/dL	0.35	0.04
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	115 mg/dL	0.26	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>