

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Nagalur  
 Lab No 12629  
 Month July  
 Year 2023  
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	108 mg/dL	-0.22	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34	6 mg/dL	0.46	0.12
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	111 mg/dL	-0.43	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

<a href="#">Home Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Lab Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

- Nagalur  
 - Valan  
 - AT...  
 - Sele...  
 ERAS  
 7.23 108  
 Sign