

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Nagalur
 Lab No 12829
 Month June
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	298 mg/dL	0.77	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.50	0.17	2.8 mg/dL	0.75	0.06
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	114 mg/dL	0.85	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links