

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Nagalur
 Lab No: 12625
 Month: May
 Year: 2023
 Constituent Group: Chemistry B

[Click on the analysis to view Graphical Data](#)

[Details About Robot Analysis](#)

[Details About Monthly Summary](#)

[Details about SDI](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 23/05/2023

Date of Report Published : 05/06/2023

Sl. No	Parameter	Method	Robotik (Semi automatic / Fully Automatic)	1834	196.88	5.19	9.36	171 mg/dL	-1.05	0.44
1	GLUCOSE B	GOD-POD B	Robotik (Semi automatic / Fully Automatic)	1834	196.88	5.19	9.36	171 mg/dL	-1.05	0.44
2	CREATININE B	Jaffes Kinetic - Alkaline Picrate B	Biosystems	38	1.75	8.31	0.16	2 mg/dL	1.38	0.05
3	CHOLESTEROL B	CHOD-PAP B	Robotik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal
Beyond ±3.0	Unacceptable performance. Action Signal

Links

[Quality Assurance](#) [Quality Methods](#) [Control Limit Review](#) [Details of Monthly Summary](#)
[Quality Summary](#) [Change Processes](#) [Notes](#) [Feedback](#)