

VIEW LAB MONTHLY SUMMARY

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Lab Name: Govt. PHC, Nuzvid  
 Lab No: 12652  
 Month: May  
 Year: 2023  
 Consultant Group: Chemistry III

[Click on the analyzer to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 23/05/2023

Date of Report Published : 05/06/2023

1	GLUCOSE III	GGD-JCO III	Robotic / Semi automatic / Fully Automatic	1834	180.88	5.19	9.39	171 mg/dL	-1.05	0.44
2	CREATININE III	Jedaa kinetic - Alkaline Phosphate III	Biosystems	38	1.78	8.91	0.16	2 mg/dL	1.38	0.05
3	CHOLESTROL III	CHOD-PAP III	Robotic / Semi automatic / Fully Automatic	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27
<b>SDI Range</b>				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal						
Beyond ±3.0				Unacceptable performance. Action Signal						

Links

- [Home Address](#)
- [Monthly Reports](#)
- [Change Password](#)
- [Submit Lab Results](#)
- [Home](#)
- [View Lab Monthly Summary](#)
- [Sign out](#)