

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Nagalur
 Lab No: 12629
 Month: April
 Year: 2023
 Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic	1832	119.45	6.04	7.21	126 mg/dL	0.91	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15	2 mg/dL	-1.23	0.05
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic	1750	111.61	5.19	5.79	120 mg/dL	1.45	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

[Master address](#)

[Master Address](#)

[Query Statement](#)

[Change password](#)

[Submit Lab Results](#)

[Home](#)

[View Lab Monthly Summary](#)

[Sign-out](#)