



AD Start Investing on Groww.

INSTALL NOW



heartbeat  
Medical Systems

CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date: 03.10.2023
	Calibration valid: 02.04.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, NAGALUR, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	
i	Manufactured by
ii	Model
iii	Value
iv	Serial No

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016  
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106  
88833 15662

No.42/1, Navaindia Road, Peelamedu,  
CBE – 04, hbmedicalsyst@gmail.com



Jump to



Vertical



Dark Mode



Compress