

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. UPHC, Vandiyur

Lab No

10809

Constituent Group Chemistry III

Date of Result Entered:

20/07/2023

PT item Lyophilized human serum based

Date of Report Published: 07/08/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	- 11
31.110						CV	SD	Value	סטו	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	108 mg/dL	-0.22	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.2 mg/dL	1.06	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	115 mg/dL	0.43	0.22

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. UPHC, Vandiyur THE MEDICAL OFFICER, NEW LIG COLONY ANNA NAGAR, VANDIYUR Madurai-Corp TAMILNADU625020

> Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Panela Christudoss

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

********* End of Report *******