



SUMMARY

Govt. PHC, Ullukurukki  
 12326  
 July  
 2023  
 Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SIM](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)  
 Date of Result Entered : 12/07/2023  
 Date of Report Published : 07/08/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
					CV	SD		
Urea Nitrogen (Urea) III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	112.1 mg/dL	0.45
Urea Nitrogen (Urea) III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.61 mg/dL	-0.63
Cholesterol (Chol) III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	111 mg/dL	-0.43

Range	Interpretation
to +1.00	Excellent.
to ±2.00	Good.
to +3.99	Accept with caution. Warning Signal.

