



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. PRIMARY HEALTH CENTRE
 Constituent Group Chemistry III
 PT item Lyophilized human serum based

Lab No 13127
 Date of Result Entered : 20/09/2023
 Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	337 mg/dL	1.75	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	116 mg/dL	0.55	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PRIMARY HEALTH CENTRE
 Govt. PHC, Kunnathur . T
 Kunnathur . T
 TAMILNADU

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

K. Christudoss
**Medical Officer,
 Govt. Primary Health Centre
 T.Kunnathur.**

Pamela Christudoss
**Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore**

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential