



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. PRIMARY HEALTH CENTRE  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 13127  
Date of Result Entered : 19/08/2023  
Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	189 mg/dL	1.49	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.68 mg/dL	-0.67	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	117 mg/dL	0.69	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PRIMARY HEALTH CENTRE  
Govt. PHC, Kunnathur . T  
Kunnathur . T  
TAMILNADU

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*K. Alleen*  
**Medical Officer,  
Govt. Primary Health Centre  
T. Kunnathur.**

*Pamela Christudoss*  
**Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore**

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential