

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Bellarympalli
 Lab No 12228
 Month July
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

Sl. No.	Test Name	Unit	Result	Reference Range	CV	CV	CV	CV
1	GLUCOSE III	GOD-POD III	109.35	5.55	6.07	110 mg/dL	0.11	0.78
2	CREATININE III	Jaffe's Kinetic - Alkaline Phosphate III	5.83	8.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	113.00	4.12	4.66	109 mg/dL	-0.88	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within +1.01 to +2.00	Good
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance Action Signal.

Links

Home	Lab Test	Lab Test	Lab Test
Home	Lab Test	Lab Test	Lab Test