



ELECTRICAL SAFETY TESTING/CALIBRATION	
Report No : HBMS/CAL/23-24	Calibration Date : 30.09.2023
	Calibration Valid : 29.03.2024
1.1 CUSTOMER DETAILS:	
Name and Address of the organization	M/S, URBAN PRIMARY HEALTH CENTRE, PATTUNOOL.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	SEMI AUTO ANALYZER
i	Manufactured by MEDSORCE
ii	Model ---
iii	Serial No 6018001218
iv	Location LABORATORY

2. CONDITION OF THE ITEM WHEN RECEIVED
No visible damages and in working order

3. TEST DONE
Electrical safety/ Performance Testing

4. STATUS						
Manufactures Specification	User specification	Within specification	Out of specification	Calibration	Electrical Safety Test	Performance Analysis
√	-	√	-	-	√	-

MD Reg No: TN/CBE/MD42/00016
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106
88833 15662

No.42/1, Navaindia Road, Peelamed
CBE - 04, hbmedicalsyster



5. TRACEABILITY DETAILS OF INSTRUMENTS USED FOR TESTING					
S.no	Name of the instrument	Make	Model	Serial no	Cal Due
1.	Electrical safety analyser	Fluke	Esa612	3005050	07-OCT-2023

6. ELECTRICAL SAFETY TEST				
S.NO	PARAMETER	OBSERVED VALUE	ACCEPTANCE LIMITS AS PER THE STD	REMARKS
1.	Protective earth resistance	0.129Ω	<0.3 Ω	Ok
2.	Earth leakage	134.21μA	<210 μA NC	Ok
3.	Patient leakage current	NA	<100 μA BF	Ok
4.	Equipment current	0.3A	As per the manufacture spec	Ok
5.	Mains voltage (P-E)	230.5V	-	Ok
6.	Mains voltage (P-N)	231.2V	-	Ok
7.	Mains voltage (N-E)	1.0V	<5V	Ok
8.	Enclosure leakage current	50 μA	<100 μA (normal)	Ok
9.	Insulation resistance	OL	>100 MΩ	Ok

S.NO	7. Remarks
1.	This report is applicable to the sample tested only
2.	The instrument used for testing are valid calibration and are traceable to national standards
3.	Parameter of the DUT were verified and found to be within the specified limits

Verified by :

 

①

②

MD Reg No: TN/CBE/MD42/00016
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106
88833 15662

No.42/1, NavalIndia Road, Peelamedu,
CBE - 04, hbmedicalsystem@gmail.com