



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13820

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
ANAKAPUTHUR-600070, CHENGALPATTU DISTRICT.

SRF No.	SRF/23/00819-0002
SRF Date	06-10-2023
Date of Receipt	06-10-2023
Date of Calibration	06-10-2023
Due Date for Calibration	05-10-2024
Issue Date	07-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW05398	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.6	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10	9.96	9.95	9.98	9.96	9.97	9.97	-0.03	0.01	0.47	
		9.98	9.96	9.97	9.98	9.95					
2	50	49.86	49.85	49.87	49.86	49.85	49.91	-0.09	0.05	0.47	
		49.95	49.96	49.97	49.95	49.95					
3	100	99.98	99.97	99.98	99.96	99.98	99.96	-0.04	0.01	0.47	
		99.95	99.96	99.94	99.95	99.96					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water containing Grade3 as specified in ISO 3696.

Calibrated By,


(Calibration Engineer)
M.RAGUL



Authorised by:


(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13819

Page No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
ANAKAPUTHUR-600070, CHENGALPATTU DISTRICT.

SRF No.	SRF/23/00819-0001
SRF Date	06-10-2023
Date of Receipt	06-10-2023
Date of Calibration	06-10-2023
Due Date for Calibration	05-10-2024
Issue Date	07-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW02467	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.6	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP122/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	100	99.95	99.89	99.87	99.86	99.85	99.87	-0.13	0.03	0.47	
		99.86	99.87	99.86	99.87	99.85					
2	500	499.89	499.87	499.89	499.86	499.85	499.85	-0.15	0.04	0.47	
		499.85	499.76	499.85	499.86	499.85					
3	1000	999.78	999.79	999.78	999.77	999.79	999.78	-0.22	0.01	0.47	
		999.75	999.76	999.78	999.78	999.78					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,


 (Calibration Engineer)
 M.RAGUL



Authorised by:


 Quality Manager/Chief Executive
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13891	MEDICAL DEVICES	Page No: 1 of 1
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00819-0003 & 06-10-2023	
Receipt Date	06-10-2023	
Calibration Date	06-10-2023	
Calibration Due	06-10-2024	

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
ANAKAPUTHUR-600070, CHENGALPATTU DISTRICT.

Details of Device Under Calibration (DUC)

Description : SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH
Range : MULTI	Sr. No : ATCD0401220RBK
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 25.6 °C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 86	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µA for B,BF,CF	Measured values in µA 178	Uncertainty in % (±) 7.5
3	Enclosure Leakage <500µA for B,BF,CF	Measured values in µA 224	Uncertainty in % (±) 7.2

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory


(Calibration Engineer)
E.ESWAR




Technical Manager
C. SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13892	MEDICAL DEVICES	Page No: 1 of 1
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00819-0004 & 06-10-2023	
Receipt Date	06-10-2023	
Calibration Date	06-10-2023	
Calibration Due	06-10-2024	

Customer Name & Address
 GOVERNMENT PRIMARY HEALTH CENTRE,
 ANAKAPUTHUR-600070, CHENGALPATTU DISTRICT.

Details of Device Under Calibration (DUC)			
Description	ELECTRICAL SAFETY(MICROSCOPE)	Make & Model	LABOMED & NA
Range	MULTI	Sr. No	130660620
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY


RESULTS			
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 98	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µA for B,BF,CF	Measured values in µA 204	Uncertainty in % (±) 7.3
3	Enclosure Leakage <500µA for B,BF,CF	Measured values in µA 214	Uncertainty in % (±) 7.3

- REMARKS**
1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 2. The user should determine the suitability of the instrument for its intended use.
 3. The recalibration interval should be determined on the user requirement.
 4. The results stated in this certificate relate only to the item calibrated.
 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

E. Eswar
 (Calibration Engineer)
 E.ESWAR



C. Shanmugaraj
 Technical Manager
 C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.