

VIEW LAB MONTHLY SUMMARY

|                          |                         |   |
|--------------------------|-------------------------|---|
| <b>Lab Name</b>          | Govt. PHC, Anaikudichan | <a href="#">Details About Robust Analysis</a> |
| <b>Lab No</b>            | 12001                   | <a href="#">Detail About Monthly Summary</a>  |
| <b>Month</b>             | April                   | <a href="#">Detail about SDI</a>              |
| <b>Year</b>              | 2023                    |   |
| <b>Constituent Group</b> | Chemistry III           |   |

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

| SLNo | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value | SDI  | U    |
|------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|------|------|
|      |                |                                       |  |                    |        | CV           | SD   |            |      |      |
| 1    | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1832               | 119.45 | 6.04         | 7.21 | 125 mg/dL  | 0.77 | 0.34 |
| 2    | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1820               | 2.16   | 8.09         | 0.18 | 2.4 mg/dL  | 1.37 | 0.01 |
| 3    | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1750               | 111.61 | 5.19         | 5.79 | 124 mg/dL  | 2.14 | 0.28 |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

| Links                          |                                 |                                    |  |
|--------------------------------|---------------------------------|------------------------------------|--|
| <a href="#">Modify Address</a> | <a href="#">Modify Methods</a>  | <a href="#">Submit Lab Results</a> | <a href="#">View Lab Monthly Summary</a> |
| <a href="#">Yearly Summary</a> | <a href="#">Change Password</a> | <a href="#">Home</a>               | <a href="#">Sign-out</a>                 |

