

VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. PHC, Anaikudichan	<a href="#">Details About Robust Analysis</a>
<b>Lab No</b>	12001	<a href="#">Detail About Monthly Summary</a>
<b>Month</b>	September	<a href="#">Detail about SDI</a>
<b>Year</b>	2023	
<b>Constituent Group</b>	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	320 mg/dL	0.61	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	118 mg/dL	1.01	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

<b>Links</b>			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>