

**VIEW LAB MONTHLY SUMMARY**

**Lab Name**      *Govt. PHC, Anaikudichan*

[Details About Robust Analysis](#)

**Lab No**      *12001*

[Detail About Monthly Summary](#)

**Month**      *June*

[Detail about SDI](#)

**Year**      *2023*

**Constituent Group**      *Chemistry III*

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

**Date of Result Entered :** 19/06/2023

**Date of Report Published :** 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	310 mg/dL	1.58	0.69
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.8 mg/dL	1.34	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	120 mg/dL	1.24	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

<a href="#">Links</a>			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>