



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name: Govt. PHC, Osur Lab No: 12967
Constituent Group: Chemistry III Date of Result Entered: 20/09/2023
PT Item: Lyophilized human serum based Date of Report Published: 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	312.41	4.03	12.59	284 mg/dL	-2.26	4.20
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.30	7.23	0.09	1.3 mg/dL	0.00	0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	35	113.55	3.41	3.87	112 mg/dL	-0.40	1.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Osur
OSUR PHC
Osur
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

M. L. S.
MEDICAL OFFICER
Govt. Primary Health Centre
Osur - 604 405
Vandavasi Taluk, T.V. Malai Dt.