

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Kappalavadi  
 Lab No: 12309  
 Month: September  
 Year: 2023  
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	306 mg/dL	-0.32	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.27 mg/dL	-0.19	0.00
3	CHOLESTROL. III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	114 mg/dL	0.09	0.20
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within +1.01 to +2.00				Good.						
Within +2.01 to +2.99				Accept with caution. Warning Signal.						
Beyond +3.0				Unacceptable performance. Action Signal.						

Links