

CALIBRATION CERTIFICATE

Date: 29.09.2023

Calibration Name & Address

Government Primary Health Centre

Osur

Certificate No

AWMS47C2

Details of Device Under Calibration (DUC)

Description : Micro Pipette
Range : 10 - 100ul, 100 - 1000ul
Least Count : ±2
DUC Condition

Make:
Model:
Sr.No : NW454511
Location : Lab

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 25°C	Relative humidity	49%RH
Sample Calibration date	06.03.23	Calibration Done at	ON SITE

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS			DEVIATION	EXPANDED UNCERTAINTY(±)2ul
		100ul	1000ul	99ul		
1	100 - 1000ul	100ul	1000ul	99ul	-2ul	OK
2	10 - 100ul	100ul	99ul	100ul	-1ul	OK

Remarks

- 1.This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2.The user should be determine the suitability of the instrument for its intended use
- 3.The Recalibration interval should be determine on the User requirement.
- 4.The results Stated in this Certificate relate only to the Item Calibrated
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00
- 6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By

