

## VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Poovaimanagar		<a href="#">Back to Search Page</a>							
Lab No	12528									
Year	2023									
Month	September		<a href="#">Print</a>							
Parameter Group	Chemistry III									
Result Entered Date : 20/09/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	312.41	4.03	12.59			4.20
			Transasia / Erba	18	307.20	6.79	20.85			9.83
			Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	320 mg/dL	0.61	0.69
			Any Analyser (Automation / Semi Automation )	68	312.35	4.82	15.05			3.65
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.30	7.23	0.09			0.03
			Transasia / Erba	21	1.26	9.52	0.12			0.05
			Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.33 mg/dL	0.38	0.00
			Any Analyser (Automation / Semi Automation )	69	1.31	8.62	0.11			0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	35	113.55	3.41	3.87			1.31
			Transasia / Erba	19	113.64	2.93	3.33			1.53
			Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32			0.20
			Any Analyser (Automation / Semi Automation )	151	113.39	4.43	5.02	118 mg/dL	0.92	0.82
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)