

## VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Poovaimanagar		<a href="#">Back to Search Page</a>							
Lab No	12528									
Year	2023									
Month	July		<a href="#">Print</a>							
Parameter Group	Chemistry III									
Result Entered Date : 19/07/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	108.24	4.71	5.10			1.70
			Transasia / Erba	24	109.83	6.83	7.50			3.06
			Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	116 mg/dL	1.10	0.28
			Any Analyser (Automation / Semi Automation )	75	109.71	5.43	5.95			1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	5.84	5.91	0.34			0.12
			Transasia / Erba	18	5.79	3.00	0.17			0.08
			Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.4 mg/dL	-1.23	0.02
			Any Analyser (Automation / Semi Automation )	66	5.89	4.60	0.27			0.07
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	113.73	3.56	4.04			1.39
			Transasia / Erba	19	112.64	4.38	4.93			2.26
			Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66			0.22
			Any Analyser (Automation / Semi Automation )	148	112.78	4.33	4.88	112 mg/dL	-0.16	0.80
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)