

## VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Poovaimanagar		<a href="#">Back to Search Page</a>							
Lab No	12528									
Year	2023									
Month	June		<a href="#">Print</a>							
Parameter Group	Chemistry III									
Result Entered Date : 20/06/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	280.51	4.07	11.42			3.81
			Transasia / Erba	22	283.77	5.80	16.44			7.01
			Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	278 mg/dL	-0.58	0.69
			Any Analyser (Automation / Semi Automation )	65	287.55	4.50	12.95			3.21
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.50	0.17			0.06
			Transasia / Erba	23	2.50	9.65	0.24			0.10
			Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.45 mg/dL	-0.70	0.01
			Any Analyser (Automation / Semi Automation )	67	2.58	6.39	0.16			0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	113.79	4.23	4.81			1.58
			Transasia / Erba	22	114.61	5.43	6.22			2.65
			Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07			0.24
			Any Analyser (Automation / Semi Automation )	149	114.61	4.05	4.64	118 mg/dL	0.73	0.76
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Home](#)