



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO: SBS/CL/23/14352</b>		<b>MEDICAL DEVICES</b>		<b>Page No:1 of 1</b>
Issue Date		14-10-2023		
SRF No & Date		SRF/23/00880-0001 & 13-10-2023		
Receipt Date		13-10-2023		
Calibration Date		13-10-2023		
Calibration Due		12-10-2024		
<b>Customer Name &amp; Address</b>				
GOVERNMENT PRIMARY HEALTH CENTRE, KAPPALVADI-635104, KRISHNAGIRI DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description : SEMI AUTO ANALYZER		Make & Model : ROBONIK & PRIETEST TOUCH		
Range : MULTI		Sr. No : ATP0381021RBK		
Resolution : MULTI		Identification No : NA		
DUC Condition : SATISFACTORY		Location : LABORATORY		
<b>Environmental Conditions &amp; Calibration Procedure Details</b>				
Environmental Details		Temperature:25.6°C		Relative Humidity 52% RH
Calibration Procedure No		SBS/CP/MD/20		Calibration done at ONSITE
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024


### ELECTRICAL SAFETY

#### RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 87	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 195	Uncertainty in % (±) 9.3
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 241	Uncertainty in % (±) 8.4

#### REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the Instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By  
  
 (Calibration Engineer)  
 E.ESWAR



Authorised Signatory  
  
 Technical Manager  
 (C.SHANMUGARAJ)

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.