

HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),
IVAKASI, VIRUDHUNAGAR-626124.
E-mail : hopeultrasystems@gmail.com

GSTIN. 33DLMPK3057D1Z9
Mobile: +919487402761, +919715223192.

CALIBRATION CERTIFICATE

Requested By

GPHC, MARANERI.

Scope:

Equipment Name : Micropipette
Make : Microlux
Model : 100-1000 μ l
SI no :

Asset no : PHC/MN-LAB/001
Calibration Done at : customer place
Calibration Date : 31/08/2023
Next call due on : 30/08/2024

TEST RESULTS Volume μ l

Sl. No.	DUC reading(μ l)	Standard Reading (μ l)		Correction (μ l)
		Trail 1	Trail 2	
01	100	100.0	100.0	+0
02	200	200.0	200.0	+0
03	500	500.0	500.0	+0
04	800	800.0	800.0	+0
05	1000	1000.0	1000.0	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

1. Electronic weighing scale(EWC/SF400C)

Note:

1. Calibration points were selected as per customer request.
2. Each reading was taken by repeatability.
3. Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems



Authorized signatory