

		CAL	IBRAT	TION CER	TIFICA	<u>TE</u>	
CERTIFICATE NO:	STIFICATE NO: SBS/CL/23/14166			MECHANICAL			Page No:1 of 1
Issue Date				12-10-2023		222	
SRF No & Date				SRF/23/00861-0002 & 11-10-2023			
Receipt Date			19 15 Th 1956 CH.	11-10-2023			
Calibration Date			18 18 19 19 19 19 19 19 19 19 19 19 19 19 19	11-10-2023			
Calibration Due				10-10-2024			
Customer Name & A	ddress						
GOVERNMENT PRIM	IARY HEALTI	CENTRE					
NACHIKUPPAM-6351	21,KRISHNA	GIRI DIST	RICT.				
		D	etails of De	vice Under Calib	ration (DUC)	
escription CENTRIFUGE				Make & Model :		BD INSTRUMENTATIONS & BDI-152	
Range :	350	0 RPM	1	Sr. No	:	NA	
Resolution :		1 RPM	1	dentification No		NA	
OUC Condition :	Satisfact	orv	L	ocation	:	LABORATORY	
	Envi	ronmental	Conditions	s & Standard Op	erating Proc	edure Details	
invironmental Details Temperature					55% Rh		
alibration Procedure No SBS/CP/N		/L/04 Calibration		done at ONSITE			
			Refere	ence Standards D	etails		
No Description	Description Ma		Make/ SI	No:	Certificate No		Validity
Digital Tachometer			LINE SEI	LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013	
ALIBRATION RESUL							

DEVIATION

RPM

1.2

0.4

0.4

-

S.No

3

DEVICE UNDER

CALIBRATION

RPM

1000

2000

3500

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

STANDARD

INSTRUMENTS

RPM

998.8

1999.6

3499.6

- 2. The user should determine the suitability of the instrument for its intended use.
- The user should determine the suitability of the instrument for its intended used.
 The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

M

(Calibration Engineer)

M.DINESH



Authorised Signatory
(Quality Manager)
(C.SIVABALAN)
(Chief Executive)

EXPANDED UNCERTAINTY

0.1

5.8

5.8

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration @ sunshinebiomedical.com, Website: www.sunshinebiomedical.com.