



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

12975

Lab Name Govt. UPHC, Uphc Vandavasi

Lab No

20/09/2023

Constituent Group

Chemistry III

Date of Result Entered :

05/10/2023

PT Item Lyophilized human serum based

Date of Report Published :

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
-------	---------	-------------------------	---------------	--------------------	----	----	----	------------	-----	---

1	GLUCOSE III	GOD-POD III	Robonik (Semi Automatic)	1870	310.86	4.82	14.98	293 mg/dL	-1.19	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi Automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.42 mg/dL	1.24	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi Automatic / Fully Automatic)	1783	113.62	3.80	4.32	109 mg/dL	-1.07	0.20

SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

LAB ADDRESS :

Govt. UPHC, Uphc Vandavasi

URBAN PHC

Uphc Vandavasi

TAMILNADU

Pamela Christodoss

Dr. Pamela Christodoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Coordinator Contact Details:

Email:clifnq@cmcvellore.ac.in

Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

C. Ramesh Babu