

ALLWIN MEDICAL SYSTEMS

No. 2/4, 2nd Street, 1st Floor, Jayalakshmi Nagar, Kattupakkam, Chennai - 600 056, Tamilnadu, India,

Cell: 9443663366

Email: allwinmedicalsystems@gmail.com

CALIBRATION CERTIFICATE

Date: 29.09.2023 AWMS73C23

Calibration Name & Address

Urban Primary Health Centre

Vandavasi

Details of Device Under Calibration (DUC)

Description Range

: Centrifuge : 0-3000 RPM Make : Remi Model : R - 6C

Least Count

: ±10 RPM

Sr.No:

DUC Condition

Location : Lab

Environmental Conditions & Calibration Procedure Details

Environmental Details

CALIBRATION RESULTS

Temperature: 25°C

Relative humidity

Certificate No

51%RH

Ok

Sample Calibration date 06.03.23

3000

Calibration Done at ON SITE

Ok

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS	DEVIATION	EXPANDED UNCERTAINTY(±) 25 RPM
		RPM		
1	1200	1200	Ok	Ok
2	1800	1800	Ok	Ok
3	2400	2400	Ok	Ok

Remarks

1. This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory

3000

- 2. The user Should be determine the suitability of the instrument for its intended use
- 3. The Recalibration interval should be determine on the User requirement.
- 4. The results Stated in this Certificate relate only to the item Calibrated
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00

6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By

Authorized Signatory

S. may