



ALLWIN MEDICAL SYSTEMS

No. 2/4, 2nd Street, 1st Floor, Jayalakshmi Nagar,
Kattupakkam, Chennai - 600 058, Tamilnadu, India,

Cell : 9443663366

Email : allwinmedicalsysteMS@gmail.com

CALIBRATION CERTIFICATE

Date: 29.09.2023

Calibration Name & Address		Certificate No	AWMS71C23	
Urban Primary Health Centre Vandavasi				
Details of Device Under Calibration (DUC)				
Description	: Micro Pipette	Make:	Thermoscientific	
Range	: 0 - 10ul, 100 - 1000ul	Model :	F3	
Least Count	: ±2	Sr.No :	RW10181, RW12536	
DUC Condition	:	Location :	Lab	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:	Relative humidity	49%RH	
Sample Calibration date	06.03.23	Calibration Done at	ON SITE	

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS			DEVIATION	EXPANDED UNCERTAINTY(±)2ul
ul						
1	100 - 1000ul	1000ul	1000ul	998ul	-2ul	Ok
2	0 - 10ul	10ul	10ul	10ul	Ok	Ok

Remarks

- 1.This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2.The user Should be determine the suitability of the instrument for its intended use
- 3.The Recalibration interval should be determine on the User requirement.
- 4.The results Stated in this Certificate relate only to the Item Calibrated
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00
- 6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By

Authorized Signatory