



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name

Govt. PHC, Kavanur

Constituent Group

Chemistry III

PT Item

Lyophilized human serum based

Lab No

12552

Date of Result Entered :

17/03/2023

Date of Report Published :

04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	85.6 mg/dL	-0.86	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.36 mg/dL	0.25	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	115 mg/dL	0.70	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Kavanur

Kavanur

TAMILNADU

Coordinator Contact Details:  
Email: [clinqcc@cmcvellore.ac.in](mailto:clinqcc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Paula Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*Myra Ramesh*  
Assistant Surgeon  
Govt. Primary Health Center  
R. Kavanur,  
Ramanathapuram, Dist.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*