

**MONTHLY SUMMARY**

Govt. PHC, Ullukurukki  
 12326  
 May  
 2023  
 Patient Group: Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/05/2023  
 Date of Report Published : 05/06/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	186.4 mg/dL	0.59	0.44
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.83 mg/dL	0.53	0.01
CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	116.5 mg/dL	0.33	0.27

SDI Range	Interpretation
> +1.00 to +1.00	Excellent.
> +1.01 to +2.00	Good.
> +2.01 to +2.99	Accept with caution. Warning Signal