



Govt. PHC, Ullukurukki
 12326
 June
 2023
 Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	289.1 mg/dL	0.17	0.69
III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.47 mg/dL	-0.58	0.01
DL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	112 mg/dL	-0.34	0.24

Range	Interpretation
+1.00	Excellent.
±2.00	Good.
+2.99	Accept with caution Warning Signal