



MARY

Govt. PHC, Ullukurukki

12326

August

2023

Chemistry III

Click on the analyte to view Graphical Data

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

e	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	189 mg/dL	1.49	0.42
I	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.77 mg/dL	0.08	0.01
III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	108 mg/dL	-1.25	0.22

ge	Interpretation
1.00	Excellent.
2.00	Good.
2.99	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

Activate Windows
Go to PC settings to activate Windows.