

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Chinthamani

[Details About Robust Analysis](#)

Lab No 10769

[Detail About Monthly Summary](#)

Month June

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	269 mg/dL	-1.19	0.69
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	114 mg/dL	0.05	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>