

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Vellodu.R  
 Lab No: 12156  
 Month: August  
 Year: 2023  
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

Details About Robot Analysis  
 Detail About Monthly Summary  
 Detail about SDI

All Analyser Result Print

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl No	Analyte	Method / Analyser used	Analyser Name	No of Participants	CV	CV	SD	Mean Value	SD	CV
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	162 mg/dL	-1.40	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	119 mg/dL	1.12	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

Address	Mobile No/No	Submit Lab Results	View Lab Monthly Summary
Secretary	Change Password	Home	Sign-out

மருத்துவ அலுவலர்  
 அரசு ஆரம்ப சுகாதார நிலையம்,  
 R.வெள்ளோடு