



MUTHU SURGICALS

101, ANNAI COMPLX, NEW BUS STAND
PERAMBALUR-621212

UDYAM REG NO : TN-16-000072

GST NO : 33DDEPM1487D1ZQ

PAN NO : DDEPM1487D

CALIBRATION CERTIFICATE

CERTIFICATE NO: MS1251072023

MECHANICAL

Page No.1 of 1

Issue Date

Receipt Date

Calibration Date

Calibration Due

01-09-2023

04-09-2023

04-09-2023

05-09-2024

Customer Name & Address

THE MEDICAL OFFICER,
POYYUR PHC,
621707,

Details of Device Under Calibration (DUC)

Description

Range

Least Count

DUC Condition

CENTRIFUGE

3500 RPM

1 RPM

Satisfactory

Make & Model

Sr. No

Identification

Location

REMI \ BDI-158

12965

NO

LAB

Environmental Details

Environmental Conditions & Standard Operating Procedure Details

Calibration Procedure No

Temperature: 26.2 °c

Relative Humidity

48% Rh

Reference Standards Details

Calibration done at

ONSITE

S No

Description

Make/ SI No

Certificate No

Validity

1

Digital Tachometer

LINE SEIKI / 175-0034V

TMS/21/257 02

05-09-2024

CALIBRATION RESULT :

No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
1.	RPM	RPM	RPM	(1) %
	3000	2995.4	-@ 6	10

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement
4. The results stated in this certificate relate only to the item calibrated
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
6. Equipment used for Calibration were calibrated & traceable to National & International Standards

Calibrated By
(BIO MEDICAL ENGINEER)
(N.MUTHAZHAGAN BE.,(BME))

Authorized Signatory

Chief Executive

For MUTHU SURGICALS

PROPRIETOR