




CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, KOLLAMCODE-629 160.		Certificate No:	VCPL/CC/23-24/47-002		
		Date of Calibration:	14-10-2023		
		Date of issue:	14-10-2023		
		Date of Receipt:	14-10-2023		
		Next cal Date:	13-10-2024		
Details of Device Under Calibration					
Description	MICROPIPETTE		Make:	THERMO SCIENTIFIC	
Range	(100 to 1000) μ l		Model/Type	FINNPIPETTE F3	
Resolution	0.2 μ l		Condition of DUC	GOOD	
Serial Number	RW02682		Location:	LABORATORY	
ID number	-		Cal. At:	LAB	
Environmental Conditions			Calibration Method Used		
Temperature	23.5 °C		Calibration Method Used	ISO 8655:2002 PART-6	
Humidity	51 % Rh		Cal procedure no:	-	
Master Details					
SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024
RESULT OF CALIBRATION					
S.NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)		ERROR (mg)	
1	100.0	99.89		0.11	
2	500.0	499.84		0.16	
3	1000.0	999.79		0.21	

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

 C.AKASH
 (CALIBRATION ENGINEER)



Authorised by:

 M. BALAJI,
 Quality Manager
 Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd
 (A Unit of Rainbow Bio Medical Engineers)
 222, East Nandan Apartments, Indira Nagar, 250 Raddiyar Main Road, Marasamangudi, Chennai - 600 044.
 No. 85/07/08, Anandarama Murugan Temple, 2nd Floor, KVM Plaza, 2nd Cross Street, Thanganagar, Pallicherry, Chennai - 600 004.
 Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376
 ■ vanavilcalibrations@gmail.com ■ www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com

 Select

 Done
