



### CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT PRIMARY HEALTH CENTRE, KOLLAMCODE-629 160.		Certificate No:	VCPL/CC/23-24/47-003		
		Date of Calibration:	14-10-2023		
		Date of issue:	14-10-2023		
		Date of Receipt:	14-10-2023		
		Next cal Date:	13-10-2024		
<b>Details of Device Under Calibration</b>					
Description:	MICROPIPETTE	Make:	THERMO SCIENTIFIC		
Range:	( 10 to 100 ) $\mu$ l	Model/Type:	FINNPIPETTE F3		
Resolution:	0.2 $\mu$ l	Condition of DUC:	GOOD		
Serial Number:	RW01332	Location:	LABORATORY		
ID number:	-	Cal. At:	LAB		
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>			
Temperature:	23.5 °C	Calibration Method Used:	ISO 8655:2002 PART-6		
Humidity:	51 % Rh	Cal procedure no:	-		
<b>Master Details</b>					
SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024
<b>RESULT OF CALIBRATION</b>					
S.NO	Duc Readings ( $\mu$ l )	Standard Reading Mean Value ( mg )	ERROR (mg)		
1	10.0	9.89	0.11		
2	50.0	49.84	0.16		
3	100.0	99.72	0.28		


**Remarks:**

- 1.DUC is Defined as Device Under Calibration.
2. Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

C.AKASH

(CALIBRATION ENGINEER)



Authorised by:

M. BALAJI,  
Quality Manager  
Vanavil Calibrations Pvt Ltd.

### Vanavil Calibrations Pvt. Ltd

A Unit of Rainbow Bio Medical Engineers

D, Multi Storey Apartment, Indira Nagar, 250 Reddyar Main Road, Narasimhapeta, Salem - 636 002

No. 25/27 St. Radhakrishnan Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangan Nagar, Palakkad, District Palakkad

Call: 99949 13872 / 99949 23872. Ph: 0427-7965876

vanavilcalibrations@gmail.com
 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com

 Select

Done

