



CALIBRATION CERTIFICATE

Page : 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE , KOLLAMCODE-629 160.	Certificate No:	VCPL/CC/23-24/47-001
	Date of Calibration:	14-10-2023
	Date of issue:	14-10-2023
	Date of Receipt:	14-10-2023
	Next cal Date:	13-10-2024

Details of Device Under Calibration

Description	CENTRIFUGE	Make:	DALAL
Range	5000 RPM	Model/Type:	MCD-4-M
Resolution	10 RPM	Condition of DUC:	GOOD
Serial No / Id.No	GPHC/LAB/CEN/01	Instrument Location:	LABORATORY

Environmental Conditions

Temperature	23.4 °C	Reference Standard No	SANAS TR4501
Humidity	52 % Rh	Calibration Procedure No	-

Method Used For Calibration

Details of Reference Standard Equipment For Calibration

S.No	Description	ID No / SL No	Certificate No	Validity
1	DIGITAL TACHOMETER	DTM-01	HT/CC/221202-03/005 SYSTEMS/ HTM 560	4-Dec-23

Calibration Result

S.No	DUC Reading (RPM)	Standard Reading (RPM)	Error (RPM)	Measurement Uncertainty (RPM)
1	500	499.8	0.2	12
2	1000	1001	-1	12
3	2000	1999	1	12
4	4000	3998	2	12
5	5000	4997	3	12

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

C.Akash
C.AKASH
 (CALIBRATION ENGINEER)



*** End of the Report ***

Authorised by:

M. Balaji
M. BALAJI,
 Quality Manager
 Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd
 (A Unit of Rainbow Bio Medical Engineers)
 57, Riski Medical Apartment, Indira Nagar, 250 Baddiyur Main Road, Narasipet, Srisailam - 508 001
 2nd, 3rd, 4th, 5th, 6th, 7th, 8th, 9th, 10th, 11th, 12th Floor, KVR Plaza, 3rd Cross Street, Thangam Nagar, Palurapatti, Chennai - 600 016
 Cell: 99949 13672 / 99949 23672 Ph: 0427-7965376
 ■ vanavilcalibrations@gmail.com ● www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com

 Select

Done

