

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Subramaniyapuram	Details About Robust Analysis									
Lab No	10781	Detail About Monthly Summary									
Month	May	Detail about SDI									
Year	2023										
Constituent Group	Chemistry III										
		Click on the analyte to view Graphical Data				All Analyser Result		Print			
		Date of Result Entered : 22/05/2023									
		Date of Report Published : 05/06/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	199 mg/dL	1.93	0.44	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.3 mg/dL	-3.48	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out