



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. LRHC, Puthupattu

[Details About Robot Analyser](#)

Lab No: 10799

[Details About Monthly Summary](#)

Month: April

[Details about SDI](#)

Year: 2023

Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered: 19/04/2023

Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GDH-POD II	Robork ( Semi automatic / Fully Automatic )	1822	119.45	0.94	7.21	138 mg/dL	2.57	0.94
2	UREA III	Others ( Urasee Berthelol / Nessler Dye ) II	dry Analyser (Automation / Semi Automation)	96	43.27	11.30	4.80	35 mg/dL	-1.51	1.28
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate II	Robork ( Semi automatic / Fully Automatic )	1820	2.16	6.09	0.18	1.3 mg/dL	-1.49	0.91
4	CHOLESTROL III	GHOD-PAP II	Robork ( Semi automatic / Fully Automatic )	1790	111.61	0.19	0.79	119 mg/dL	1.28	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>