

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	<b>Govt. UPHC, Puttuthoppu</b>	<a href="#">Details About Robust Analysis</a>
<b>Lab No</b>	<b>10799</b>	<a href="#">Detail About Monthly Summary</a>
<b>Month</b>	<b>June</b>	<a href="#">Detail about SDI</a>
<b>Year</b>	<b>2023</b>	
<b>Constituent Group</b>	<b>Chemistry III</b>	
<a href="#">Click on the analyte to view Graphical Data</a>		<input type="button" value="All Analyser Result"/> <input type="button" value="Print"/>
Date of Result Entered : <b>20/06/2023</b>		
Date of Report Published : <b>04/07/2023</b>		

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	<b>286</b> mg/dL	-0.04	0.69
2	<a href="#">UREA III</a>	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation )	50	33.99	14.82	5.04	<b>38</b> mg/dL	0.80	1.42
3	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	<b>3</b> mg/dL	2.51	0.01
4	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	<b>117</b> mg/dL	0.64	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>