

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	Govt. UPHC, Fathima Nagar		<a href="#">Details About Robust Analysis</a>							
<b>Lab No</b>	10797		<a href="#">Detail About Monthly Summary</a>							
<b>Month</b>	July		<a href="#">Detail about SDI</a>							
<b>Year</b>	2023									
<b>Constituent Group</b>	Chemistry III									
<a href="#">Click on the analyte to view Graphical Data</a>			<a href="#">All Analyser Result</a>		<a href="#">Print</a>					
			Date of Result Entered : 20/07/2023							
			Date of Report Published : 07/08/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	107 mg/dL	-0.39	0.28
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.3 mg/dL	1.34	0.02
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	116 mg/dL	0.64	0.22
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>