



CALIBRATION / TESTING REPORT			
CUSTOMER DETAILS :  GOVT.PRIMARY HEALTH CENTRE VELAYUTHAPATTINAM, SIVAGANGAI.		CERTIFICATE NO	TICS/2023/VYP/SG/02
		EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	LABOTECH
		MODEL	--
CALIBRATION DATE	20/09/2023	ASSET NO	VYP002
CALIBRATION DUE DATE	19/09/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required  Yes  No
- 2. Any Physical damage observed  Yes  No
- 3. Spare Replacement (if required)  Yes  No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MECO	KM2234 BL	S1049875
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011

TEST REPORT			Temperature 25°C Humidity RH 55%		
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS

ROTATION PER MINUTE	RPM	Speed 2	2088	± 5	PASS
		Speed 4	2891		

\*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE  CORRECTIVE MAINTENANCE REQUIRED  REMOVED FROM USE

Calibrated By : 

Authorized Sign With Seal 

