

<u>c</u>	ALIBRAT	TON CER	TIFICA	<u>TE</u>	
CERTIFICATE NO: SBS/CL/23/14185		MEDICAL DEVICES		Page No:1 of 1	
Issue Date		12-10-2023			· · · · · · · · · · · · · · · · · · ·
SRF No & Date		SRF/23/00864-0	0001 & 11-10	-2023	
Receipt Date		11-10-2023			
Calibration Date		11-10-2023			
Calibration Due		10-10-2024			
Customer Name & Address				,, <u>=</u> ,	···
GOVERNMENT PRIMARY HEALTH CE	NTRE,				
NAGAMANGALAM-635113, KRISHNAG	IRI DISTRICT.				
	Details of De	evice Under Calibr	ation (DUC)		
Description : SEMI AUTO A	SEMI AUTO ANALYZER		Make & Model ROBONIK & PRIE		TEST TOUCH
Range : MULTI	MULTI		:	ATCD3660321RBk	(
Resolution : MULTI		Identification No	:	NA	
DUC Condition : SATISFACTOR	SATISFACTORY		:	LABORATORY	
En	vironmental Cond	itions & Calibratio	n Procedure	Details	
Environmental Details Temperature:25.6 ° C		Relative Humidity 52% RH		52% RH	·
Calibration Procedure No SBS/CP/MD/20		Calibration done at ONSITE		· · · · · · · · · · · · · · · · · · ·	
	Refer	ence Standards De	etails		
S.No Description Make/ SI N		Certificate		No	Validity
1 Electrical Safety Analyser Rigel Medic		al & 44L-1059	4L-1059 M-230809-16-4		10-08-202

ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	95	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	162	7.6
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
ľ	<500µAfor B,BF,CF	234	7.2

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards. Calibrated By **Authorised Signatory**

M.DINESH





Chief Executive