



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/00510

Page: 1 of 2

Customer Name:	GOVT COMMUNITY HEALTH CENTRE	Customer Address:	KADUGUR, ARIYUR DISTRICT.
Location:	LABORATORY	Receipt Date:	30.01.2021
Description:	MICROSCOPE	Date of Calibration:	30-01-2021
Make & Model:	ZOOM& ZBCM-08	Date of Due:	29-01-2022
Serial No:	474431	Issue Date:	02.02.2021
SRF No:	SBS/SRF/21-22/0008-014	SRF Date:	30.01.2021

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1	Electrical Safety Analyzer	Fluke	ESA615	2244202	26-Jun-2021

Environmental Conditions and Standard Operating Procedure Details:


Temperature	24°C	Calibration Done At	ONSITE
Humidity (40-70% RH)	55%	Calibration Procedure No	SBS-MD-22

REMARKS:

Electrical Safety Test Passed as Per IEC Standard 60601-1/62353

The above mentioned instruments has been calibrated using standard manufacturer recommended protocols, using equipments having traceability to National/International Standards. Partial reproduction of the certificates is not permitted.

Calibrated by:


S. LAKSHMANAN MURUGAN
(Calibration Engineer)

Approved by:


C. SHANMUGARAJ
(Technical Manager)