



## CALIBRATION REPORT

### Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/00503

Page: 1 of 2

Customer Name:	GOVT COMMUNITY HEALTH CENTRE	Customer Address:	KADUGUR, ARIYUR DISTRICT.
Location:	LABORATORY	Receipt Date:	30.01.2021
Description:	BIO-CHEMISTRY ANALYSER	Date of Calibration:	30-01-2021
Make & Model:	ROBONIK & PRIETEST TOUCH	Date of Due:	29-01-2022
Serial No:	AT0961010RBK	Issue Date:	02-02-2021
SRF No:	SBS/SRF/21-22/0008-007	SRF Date:	30.01.2021

### Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1	Electrical Safety Analyzer	Fluke	ESA615	2244202	26-Jun-2021

### Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	ONSITE
Humidity (40-70% RH)	55%	Calibration Procedure No	SBS-MD-22

### REMARKS:

Electrical Safety Test Passed as Per IEC Standard 60601-1/62353

The above mentioned instruments has been calibrated using standard manufacturer recommended protocols, using equipments having traceability to National/International Standards. Partial reproduction of the certificates is not permitted.

Calibrated by:

M. BALAJI

(Calibration Engineer)

Approved by:

  
C. SHANMUGARAJ

(Technical Manager)