



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/14819

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE(NAVALPATTU BLOCK),
KEELAKURICH-620011,TIRUCHIRAPALLI DISTRICT.

SRF No.	SRF23/00894-0004
SRF Date	14-10-2023
Date of Receipt	13-10-2023
Date of Calibration	14-10-2023
Due Date for Calibration	13-10-2024
Issue Date	16-10-2023

Details of Unit Under Calibration	
Description	MICRO PIPETTE
Range	10-100µl
Resolution	0.2µl
Serial Number	NA
ID Number	NA
Cal. AI	Mechanical Lab
Make	THERMO SCIENTIFIC
Model	FINPIPEPETTE F3
Material	PVC
Operating Range	10-100µl
Condition of UUC	Good
Instrument Location	LABORATORY

Environmental Condition		Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6
Standard Used		National / International Standard	ISO 8655-6:2002
		Cal Procedure No	SBS/CP/ML/08

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPT/22/12/2115-01	A&D & CH-252	National Standards	09-12-2023

Z Factor: 1.00319

SI. No.	Nominal Value	Result of Calibration in µl						Measurement Uncertainty (±)		
		Observed Readings			Mean Value	Systematic Error	Random Error			
1	10.0	9.91	9.92	9.90	9.92	9.93	9.92	-0.08	0.01	0.47
		9.92	9.93	9.92	9.93	9.91				
2	50.0	49.95	49.96	49.95	49.95	49.96	49.96	-0.04	0.01	0.47
		49.96	49.97	49.98	49.97	49.96				
3	100.0	99.89	99.88	99.89	99.87	99.88	99.88	-0.13	0.01	0.47
		99.88	99.87	99.85	99.86	99.87				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,
M. P. Velupillay
(Calibration Engineer)



(Signature)

Authorised by:
(Signature)



CALIBRATION CERTIFICATE

Certificate No. SBS/CL/23/4818

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE (NAVALPATTU BLOCK),
KEELAKURUCHI-620011, TRICHURAPALLU DISTRICT

SRF No.	SRF/23/06/94-0003
SRF Date	14-10-2023
Date of Receipt	13-10-2023
Date of Calibration	14-10-2023
Due Date for Calibration	13-10-2024
Issue Date	16-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	N/A	Operating Range	100-1000µl
ID Number	N/A	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & G14-252	National Standards	06-12-2023

Result of Calibration in µl

SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
		99.85	99.86	99.87	99.85	99.84				
1	100	99.85	99.86	99.87	99.85	99.84	99.84	-0.16	0.02	0.47
		99.83	99.82	99.85	99.81	99.82				
2	500	499.88	499.87	499.86	499.85	499.84	499.84	-0.15	0.02	0.47
		499.85	499.87	499.85	499.84	499.83				
3	1000	999.95	999.97	999.95	999.94	999.93	999.94	-0.06	0.02	0.47
		999.92	999.91	999.93	999.94	999.95				

Z Factor: 1.00319

Remarks

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the Laboratory.
- 2 The user should determine the suitability of the instrument for its intended use.
- 3 The recalibration interval should be determined on the user requirement.
- 4 The results stated in this certificate relate only to the item calibrated.
- 5 Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6 The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95%, for a coverage factor k=2.00
7. Calibration Liquid Used: Distilled water (Analytical Grade) as specified in ISO 3696.



Calibrated By,
M. P. Lakshmi
(Calibration Engineer)

Authorized by:
S. M. V. S.
(Quality Management Executive)



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/14816	MEDICAL DEVICES
Issue Date	16-10-2023
SRF No & Date	SRF/23/00894-0001 & 14-10-2023
Receipt Date	14-10-2023
Calibration Date	14-10-2023
Calibration Due	13-10-2024

Customer Name & Address
 GOVERNMENT PRIMARY HEALTH CENTRE(NAVALPATTU BLOCK),
 KEELAKURICHI-620011, TIRUCHIRAPALLI DISTRICT.

Details of Device Under Calibration (DUC)

Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH
Range	MULTI	Sr. No	ATCD2720818RBK
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Details	Environmental Conditions & Calibration Procedure Details
Temperature: 25.6 ° C	Relative Humidity 52% RH
SBS/CP/MD/20	Calibration done at ONSITE


Reference Standards Details	Certificate No
Make/ SI No:	M-230809-16-4
1 Electrical Safety Analyser	Rigel Medical & 44L-1059
	Validity 10-08-2024

ELECTRICAL SAFETY

RESULTS	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
S.no		Measured values in MQ	Uncertainty in % (±)
1	Insulation Resistance >20MQ	88	13.92
2	Earth Leakage <500µAfor B,BF,CF	168	7.6
3	Enclosure Leakage <500µAfor B,BF,CF	235	7.2

REMARKS

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2 The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

 (Calibration Engineer)
 E.ESWAR

Technical Manager
 C.SHANMUGARAJ

Chief Executive




 C. SHANMUGARAJ



CALIBRATION CERTIFICATE

MECHANICAL

Page No:1 of 1

CERTIFICATE NO.: SBS/CL/23/14817

16-10-2023

Issue Date

SRF/23/00894-0002 & 14-10-2023

SRF No & Date

14-10-2023

Receipt Date

14-10-2023

Calibration Date

13-10-2024

Calibration Due

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE (NAVALPATTU BLOCK),
KEELAKURICHI-620011, TIRUCHIRAPALLI DISTRICT.

Details of Device Under Calibration (DUC)

LABOTECH & BDI-152

Description : CENTRIFUGE

Make & Model

Range : 3500 RPM

Sr. No

2011543

Resolution : 1 RPM

Identification No

NA

DUC Condition : Satisfactory

Location

LABORATORY

Environmental Conditions & Standard Operating Procedure Details

Environmental Details

Relative Humidity

55% Rh

Calibration Procedure No

SBS/CP/ML/04

Calibration done at

ONSITE

S.No

Description

Make/ SI No.

LINE SEIKI / 175-0034V

Certificate No

JRPM-CCTR-A&S-2023-0013

Validity

09-06-2024

1 Digital Tachometer

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS		DEVIATION	EXPANDED UNCERTAINTY
		RPM	RPM		
1	1000	999.9	0.1	4.2	
2	2000	1999.8	0.2	4.2	
3	3000	2999.7	0.3	4.2	

REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By
E.S
(Calibration Engineer)
E.SWAR



Authorised Signatory

(Signature)
(Quality Manager)
(C.SIVABALAN)

(Signature)
(Chief Executive)

(Signature)