

Lab Name: Govt. PHC, Katpalle
Lab No: 12124
Month: September
Year: 2023
Coordinator Group: Chemistry B

[Details About Patient Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SQI](#)

Click on the analysis to view Graphical Data

Date of Result Entered: 22/09/2023
 Date of Report Published: 05/10/2023

S. No.	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	CV	Participants		Your Value	SQI	U
						CV	SD			
1	GLUCOSE (G)	GOO-FUD-G	Robotic / Semi-automate / Fully Automatic	1070	310.85	4.83	14.85	208 mg/dL	-1.00	0.00
2	CREATININE (G)	Jaffe Kinetic - Alkaline Phosphate (G)	Robotic / Semi-automate / Fully Automatic	1050	1.28	0.14	0.10	2.2 mg/dL	4.00	0.00
3	CHOLESTEROL (G)	CHOD-PAP-G	Robotic / Semi-automate / Fully Automatic	1780	113.82	0.80	0.52	100 mg/dL	-1.77	0.00

SQI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within +1.01 to +2.00	Good
Within +2.01 to +2.99	Accept with caution. Warning Signal
Beyond +3.0	Unacceptable performance. Action Signal

Home	Monthly Address	Monthly Methods	Guard Lab Results	View Lab Monthly Summary
Monthly Summary	Change Password	Home	Sign-out	

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Kalpeltal
 Lab No: 12124
 Month: February
 Year: 2023
 Consultant Group: Chemistry II

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 21/02/2023

Date of Report Published : 05/03/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robosnk (Semi automatic / Fully Automatic)	1555	155.11	6.96	10.15	155 mg/dL	4.21	0.47
2	CREATININE II	Jaffes Kinetic - Alkaline Picrate II	Robosnk (Semi automatic / Fully Automatic)	1572	3.40	9.00	0.01	3.1 mg/dL	-0.95	0.01
3	CHOLESTROL II	CHOD-PAP II	Robosnk (Semi automatic / Fully Automatic)	1755	112.79	5.23	5.90	114 mg/dL	0.20	0.25

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



Best Viewed at pixels 1024 X 768

NEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Kalpellal
 Lab No: 12124
 Month: May
 Year: 2023
 Consultant Group: Chemistry II

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)


Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GCO-PCO II	Robonik (Semi automatic / Fully Automatic)	1534	180.88	5.19	9.29	183 mg/dL	0.23	0.44
2	CREATININE II	Jaffes Kinetic - Alkaline Picrate II	Robonik (Semi automatic / Fully Automatic)	1554	1.76	7.45	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL II	CHOD-PAP II	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.63	102 mg/dL	-2.16	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by
 **Swiftware Solutions**

Best Viewed at pixels 1024 X 768

NEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Katpeltal
Lab No 12124
Month July
Year 2023
Consultant Group Chemistry II

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 15/07/2023
 Date of Report Published : 07/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOO-POD II	Robonik (Semi automatic / Fully Automatic)	1576	109.35	5.55	6.07	106 mg/dL	-0.55	0.26
2	CREATININE II	Jaffes Kinetic - Alkaline Phorate II	Robonik (Semi automatic / Fully Automatic)	1650	5.53	6.00	0.35	5.5 mg/dL	-0.09	0.02
3	CHOLESTROL II	CHOD-PAP II	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114.1 mg/dL	0.24	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Kalpettal
 Lab No: 12124
 Month: September
 Year: 2023
 Constituent Group: Chemistry II

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 22/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.06	286 mg/dL	-1.66	0.69
2	CREATININE II	Jaffes Kinetic - Alkaline Phosphate II	Robonik (Semi automatic / Fully Automatic)	1052	1.29	5.14	0.10	1.2 mg/dL	-0.66	0.00
3	CHOLESTROL II	CHOD-PAP II	Robonik (Semi automatic / Fully Automatic)	1763	113.62	3.80	4.32	106 mg/dL	-1.77	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Kalpeltal
 Lab No: 12124
 Month: June
 Year: 2023
 Consultant Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)


Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOO-POD II	Robonik (Semi automatic / Fully Automatic)	1525	266.63	5.17	14.62	275 mg/dL	-0.78	0.69
2	CREATININE II	Jaffes Kinetic - Alkaline Picrate II	Robonik (Semi automatic / Fully Automatic)	1845	2.57	6.04	0.17	2.57 mg/dL	0.00	0.01
3	CHOLESTROL II	CHOD-PAP II	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.48	8.07	109 mg/dL	-0.93	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



Best Viewed at pixels 1024 X 768