

CENTRIFUGE Calibration / Verification Record Sheet

Location: ICTC- AREA HOSPITAL- DHARMAVARAM

Range : Up to 3500 RPM

Equipment ID:APSACS/ICTC/AH/DMM/CEN/01

SL. No:02

Calibration Procedure Reference: In house verification of centrifuge using standard calibrated by Tachometer

Date of Calibration	30-09-2023	Due Date of Calibration	29-09-2024
----------------------------	------------	--------------------------------	------------

Certified Centrifuge ID: GMC-ATP/HIVTL/SRL/CEF/01	Tachometer ID: GMC-ATP/HIVTL/SRL/TM/01	Calibration date: 19-05-2023	Due date of Calibration: 18-05-2024
--	---	-------------------------------------	--

Verification Done By: P. ANURADHA, Technical Officer, SRL, GMC, Ananthapuramu

No.	STANDARD READING (RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.3	500	0.7
2.	999.6	1000	0.4
3.	1998.2	2000	1.8
4.	2997.5	3000	2.5
5.	3496.9	3500	3.1

Accept/ Reject:Accepted since average difference is within +/-1000RPM

Action Taken: NA

Remarks: NA

P. Anuradha
SRL-T.O
Dept. of Microbiology
GMC - Ananthapuramu.